

## Anti-h Collagen IV NA000706 SPTN-5

### Product overview

---

<b>Catalog number</b>	C-10-0031
<b>Specificity</b>	Antibody recognizes human Collagen IV
<b>Description</b>	Recombinant chimeric antibody containing variable region from mouse IgG and constant region from human IgG1, cultured in vitro under conditions free from animal-derived components.
<b>Product buffer solution</b>	50 mM Na-citrate, pH 6.0, 0.9 % NaCl, 0.1% ProClin™ 300 as a preservative
<b>Shelf life and storage</b>	Unspecified, storage at 2–8 °C
<b>Subclass</b>	Human IgG <sub>1</sub>
<b>Analyte description</b>	Collagen IV is a major component of basal membrane. There may contain 3 different major degradation fragments, amine terminal 7S collagen, carboxyl terminal NCL collagen and the main triple helix zone in the serum. In a normal liver, no basal membrane is present in the Disse gaps, type IV collagen concentration is very low. However, during the fibrosis process from hepatitis through to cirrhosis, the formation of basal membrane occurs in Disse gaps. Meanwhile, type IV collagen concentration in liver tissue and blood stream increases accordingly.

### Parameters tested on each lot

---

<b>Product appearance</b>	Liquid, may turn slightly opaque during storage
<b>Product concentration</b>	5.0 mg/ml (+/- 10%)
<b>Immunoreactivity</b>	80–120% compared to the reference sample
<b>IEF Profile</b>	N/D
<b>Purity</b>	≥ 95 %

### Kinetic parameters

---

<b>Association rate constant</b>	Not Determined (N/D)
<b>Dissociation rate constant</b>	N/D
<b>Affinity constant</b>	N/D
<b>Determination method</b>	-
<b>Determination antigen</b>	-



Medix Biochemica China

上海劭迪克斯生物化学有限公司

上海市闵行区浦江绿洲环路10号6幢11层

电话: 021 68119180, 68119181, 68119105

邮箱: medixchina@medixbiochemica.com

网址: www.cnmedixbiochemica.com

**Cross-reactivities** -

**Epitope** N/D

**Pair recommendations**

		DETECTION				
		NA000704	NA000705	NA000706	NA000709	NA000712
CAPTURE	NA000704	-	+	+	-	+
	NA000705	-	-	+	-	+
	NA000706	-	+	-	-	+
	NA000709	-	+	+	-	+
	NA000712	+	+	+	-	-

Please note that pair recommendations are based on results obtained by our laboratory. Equally good results may be obtained using other pairs and therefore these recommendations are only indicative.

**Platforms tested** ELISA, CLIA

**Antigens tested** -

Product stability	TEMPERATURE, TIME	RESULT
	-70 °C, 21 days	N/D
	-20 °C, 21 days	N/D
	+4 °C, 21 days	N/D
	+35 °C, 21 days	N/D
	+35 °C, 7 days	OK
	+45 °C, 7 days	OK

Stability testing is performed in the product buffer to see whether different temperatures affect the antigen binding, charge or composition of the antibody. Please note that the shelf life given on the first page is based on real time stability testing at 2–8 °C in the product buffer.

**Miscellaneous** -

**References** -

