



**Product specification** ANTIBODY

2025-10-28

## Anti-h GP73 NA001703 SPTN-5

### **Product overview**

C-10-0064 Catalog number

**Specificity** Antibody recognizes human golgi protein 73 (GP73)

**Description** Monoclonal mouse antibody, cultured in vitro under conditions free from

animal-derived components.

**Product buffer solution** 50 mM Na-citrate, pH 6.0, 0.9% NaCl, 0.1% ProClin<sup>™</sup> 300 as a preservative

Shelf life and storage Unspecified, storage at 2-8 °C

**Subclass** Mouse IgG₁

**Analyte description** GP73 is a type II Golgi transmembrane protein with a molecular weight of

> 73 KDa mainly expressed in epithelial cells. The expression of GP73 in normal liver tissue is low. But, serum GP73 detection has important clinical value for the early diagnosis of HCC and can be used as a biomarker for many other liver diseases, e.g. hepatitis C virus infection, hepatitis B virus- related cirrhosis and HCC, alcoholic hepatitis and autoimmune hepatitis, etc. GP73 can also be used in conjunction with methods such as liver stiffness measurement to evaluate the

degree of liver fibrosis and sclerosis.

#### Parameters tested on each lot

**Product appearance** Liquid, may turn slightly opaque during storage

**Product concentration** 5.0 mg/ml (+/- 10%)

80-120% compared to the reference sample in a CLIA test **Immunoreactivity** 

**IEF Profile** Not Determined (N/D)

≥ 95% **Purity** 

### Kinetic parameters

**Association rate constant** N/D

**Dissociation rate constant** N/D

**Affinity constant** N/D

**Determination method** 

**Determination antigen** 



电话: 021 68119180, 68119181, 68119105

邮箱: medixchina@medixbiochemica.com

网址:www.cnmedixbiochemica.com



# Product specification ANTIBODY

2025-10-28

Cross-reactivities

Epitope N/D

Pair recommendations

		DETECTION			
		NA001703	NA001704	NA001705	NA001706
CAPTURE	NA001703	-	+	+	+
	NA001704	+	-	+	+
	NA001705	+	+	-	-
	NA001706	+	+	-	-

Please note that pair recommendations are based on results obtained by our laboratory. Equally good results may be obtained using other pairs and therefore these recommendations are only indicative.

Platforms tested ELISA, CLIA

Antigens tested -

Product stability TEMPERATURE, TIME RESULT

-70 °C, 21 days OK
-20 °C, 21 days OK
+4 °C, 21 days OK
+35 °C, 21 days OK
+45 °C, 7 days OK

Stability testing is performed in the product buffer to see whether different temperatures affect the antigen binding, charge or composition of the antibody. Please note that the shelf life given on the first page is based on real time stability testing at 2–8 °C in the product buffer.

电话: 021 68119180, 68119181, 68119105

邮箱: medixchina@medixbiochemica.com

网址:www.cnmedixbiochemica.com

Miscellaneous -

References -

